







KAZALO:

-  BIS-MPC MMX - SOS centrala
-  BIS-NS 2405 - ispravljač 24V/5A
-  BIS-OKA 08 - osnovni komunikacijski element
-  BIS-DAC 32 - SOS centrala
-  BIS-NS 2402 - ispravljač 24V/2A
-  BIS-SS 01CR T - sobna svjetiljka s razdjelnikom
-  BIS-TR 01F - tipkalo razriješno
-  BIS-TP 02F - tipkalo pozivno krevetno
-  BIS-KS 04CB T - katna svjetiljka s biperom
-  BIS-TP SOST - tipkalo pozivno SOS

UPUTSTVO ZA INSTALACIJU:

- razvod u standardnim podžbuknim cijevima
- centrale DAC 32 se montiraju u TEM 7 mod podžbukno
- katne svjetiljke se montiraju u TEM 6 mod nadžbukno
- ispravljači se montiraju nadžbukno u kutiju 190x140x70
- sobna svjetiljka s razdjel.i tipkala montiraju se u razvodnu kutiju fi 60 mm
- tipkala se montiraju na visinu prekidača za svjetlo
- tipkalo BIS-TP SOS T se montira na visinu 200 cm
- komunikator OKA montira se u razvodnu kutiju 95 x 95
- na jedan standardni komunikator moguće je spojiti do 8 soba

BIS - SOS ADRESABILNO - ANALOGNI